

MILLER INSTRUMENTS LTD.  
 # 1 - 3871 North Fraser Way  
 Burnaby, B.C. V5J 5G6  
 Telephone (604) 431-8882  
 Fax (604) 431-8714  
 web: www.miller.bc.ca  
 email: miller@miller.bc.ca



**CERTIFICATE OF CALIBRATION**


Certificate No.: 27906  
 Date of issue: November 25, 2011  
 Manufacturer: Eurotherm Chessell (Sure Controls)  
 Model No: 6100A  
 Serial No: LH050049G  
 Description: Graphic recorder  
 Customer: Red-D-Arc Ltd.

*The Calibration Laboratory Assessment Service (CLAS) of the National Research Council of Canada (NRC) has assessed and certified specific calibration capabilities of Miller Instruments Ltd. and its traceability to the International System of Units (SI) or to the standards acceptable to the CLAS program. This certificate of calibration is issued in accordance with the conditions of certification granted by CLAS, Certification number 94-03, and the conditions of accreditation granted by the Standards Council of Canada (SCC), Accreditation number No. 156. The ISO/IEC 17025:2005 Standard was used in the above assessment carried out by CLAS.*

Date Calibrated: November 25, 2011  
 Calibration due date: November 25, 2012<sup>(4)</sup>  
 Temperature: 23±1 °C  
 Relative Humidity: 31±10 %RH

Calibrated by: *M. Mohseni*  
 (Mahkameh Mohseni, B.Sc.)

Authorized by: *W.J. Miller*  
 (S. Nishie, P. Eng., Calibration Manager)



Instrument received: <input checked="" type="checkbox"/> in-specifications* <input type="checkbox"/> out-of-specifications*	Instrument when returned: <input checked="" type="checkbox"/> meets test specifications* <input type="checkbox"/> meets limited specifications*
<input checked="" type="checkbox"/> Data supplied	<input type="checkbox"/> Data available upon request
Comments:	The instrument was calibrated after a 30-minute warm-up period.

\* The tolerance limits used in this calibration were those defined by the customer.

*For measurement results associated with the conformance to a tolerance, the uncertainty in the measurement system did not exceed 25% (4:1 test uncertainty ratio) of the acceptable tolerance for each characteristic calibrated, unless otherwise noted in the report.*

Calibration Procedure: CP-SP23439

Calibration Equipment Used:

ID #	Model	Description	Serial Number	Calibration due date
163	Fluke 5500A	Multi-Product Calibrator	8855014	Apr 7, 2012
141	Agilent 34401A	DMM	MY41002493	Oct 3, 2012

## CERTIFICATE OF CALIBRATION

Certificate No.: 27906

Page 2 of 2

Date of issue: November 25, 2011

### Calibration data (as found and as returned)

Channel No.	Applied voltage (VDC)	Expected DUT reading (°F)	DUT reading (°F)	Measurement Uncertainty (°F)	Tolerance Limits $\pm$ (°F)	Pass/fail
1	0.000	-51	-51	2	7	p
1	2.787	382	382	2	7	p
1	6.013	882	883	2	7	p
1	9.239	1382	1384	2	7	p
2	0.000	-51	-51	2	7	p
2	2.787	382	382	2	7	p
2	6.013	882	883	2	7	p
2	9.239	1382	1384	2	7	p
3	0.000	-51	-51	2	7	p
3	2.787	382	382	2	7	p
3	6.013	882	882	2	7	p
3	9.239	1382	1383	2	7	p
4	0.000	-51	-51	2	7	p
4	2.787	382	382	2	7	p
4	6.013	882	883	2	7	p
4	9.239	1382	1383	2	7	p
5	0.000	-51	-51	2	7	p
5	2.787	382	382	2	7	p
5	6.013	882	883	2	7	p
5	9.239	1382	1384	2	7	p
6	0.000	-51	-51	2	7	p
6	2.787	382	382	2	7	p
6	6.013	882	882	2	7	p
6	9.239	1382	1383	2	7	p

Note 1: DUT: Device under test.

Note 2: The DUT was powered by a 120 V AC (60 Hz) line and was calibrated after a 30-minute warm-up period.

Note 3: The uncertainty of this calibration, assuming normally distributed data, was derived from effective standard deviations and has been expanded to obtain a coverage factor of  $k=2$  at a level of confidence of approximately 95%.

Note 4: The calibration due date is shown as requested by the customer.

Note 5: The memory battery was replaced before the above calibration.