

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: *LH320164G*
Instrument Number: *8467045*
IP Address: *192.168.111.222*

Date Tested: *MAR. 2/17.*
Mac Address: *00:0A:8D:81:32:65*

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	<i>2.787</i>	<i>382</i>	<i>6.013</i>	<i>882</i>	<i>9.239</i>	<i>1382</i>
Channel 2	<i>2.788</i>	<i>382</i>	<i>6.012</i>	<i>882</i>	<i>9.240</i>	<i>1382</i>
Channel 3	<i>2.790</i>	<i>382</i>	<i>6.016</i>	<i>882</i>	<i>9.239</i>	<i>1382</i>
Channel 4	<i>2.788</i>	<i>382</i>	<i>6.012</i>	<i>882</i>	<i>9.238</i>	<i>1382</i>
Channel 5	<i>2.789</i>	<i>382</i>	<i>6.019</i>	<i>882</i>	<i>9.240</i>	<i>1382</i>
Channel 6	<i>2.789</i>	<i>382</i>	<i>6.018</i>	<i>882</i>	<i>9.238</i>	<i>1382</i>

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: *9/21/16.*

Precision digital multimeter SN: 2142011

Calibration Date: *9/21/16*

ProHeat 35 power supply used for signal output SN: *MD130155*

Calibration date: *2/16/17*

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date:

Technician: *D. George*

Signature