

# Red-D-Arc Welderrentals

## Calibration Certificate

Unit Serial Number: **LJ160028G**  
 Instrument Number: **61470**  
 IP Address: **192.168.111.222**

Date Tested: **12/08/13**  
 Mac Address: **00:0A:8D:00:F0:1E**

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	382	380	882	882	1382	1381
Channel 2	382	380	882	880	1382	1381
Channel 3	382	381	882	882	1382	1382
Channel 4	382	382	882	883	1382	1383
Channel 5	382	382	882	883	1382	1383
Channel 6	382	381	882	883	1382	1385

**Allowable difference in temp reading is +or- 5 Deg. F.**

**Instruments Used:**

Fluke 714 Thermocouple Calibrator SN: 2118055      Calibration Date: 09/08/13  
 Precision digital multimeter SN: 2142011      Calibration Date: 09/08/13  
 ProHeat 35 power supply used for signal output SN: LH360046G      Calibration date: 13/08/13

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.   
 Calibration due date: **12/08/14**

Technician: Gus Rodin  
 Signature

Montreal Canada office