

Miller ProHeat 35

Unit Serial Number: LK230182G Date Tested: 12/4/2012

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VCD@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.793	383	6.019	882	9.239	1382
TC2	2.792	383	6.012	882	9.234	1382
TC3	2.798	383	6.017	883	9.239	1382
TC4	2.800	383	6.018	883	9.245	1382
TC5	2.796	383	6.020	882	9.249	1382
TC6	2.792	383	6.013	882	9.239	1382
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1882015 Calibration Date: 12/6/11

Precision Digital Voltage Meter SN: 2036021 Calibration Date: 5/2/12

Machine re-calibration due date: 12/4/2013

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: Greg P. Date: 12/4/2012

Signature