

Red-D-Arc Welderentals

Calibration Certificate

Unit Serial Number: MA100004G

2013-Oct-08

Date Tested:

Instrument Number: 95020

Mac Address: 00:0A:8D:01:73:2C

IP Address: 192.168.111.222

Primary Standard					
382 (F)		882 (F)		1382 (F)	
Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.788	383	881	9.235	1382
Channel 2	2.792	381	882	9.237	1382
Channel 3	2.791	381	882	9.245	1383
Channel 4	2.795	383	883	9.247	1383
Channel 5	2.797	383	883	9.253	1384
Channel 6	2.792	382	883	9.250	1383

Allowable difference in temp reading is ± 0.5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118058

Calibration Date: 2013-Sep-23

Precision digital multimeter SN: 2142997


Calibration Date: 2013-Sep-23

ProHeat 35 power supply used for signal output SN: LJ220160G

Calibration date: 2103-Oct-08

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: 

Calibration due date: 2014-Oct-08

Signature