

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: MB4004456 Date Tested: 2013-Oct-18

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.795	383	6.016	883	9.245	1383
TC2	2.799	384	6.018	883	9.244	1383
TC3	2.796	384	6.019	884	9.249	1383
TC4	2.806	384	6.020	884	9.246	1384
TC5	2.798	384	6.028	884	9.250	1383
TC6	2.802	384	6.024	884	9.247	1384
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator 2118058

Calibration Date: 2013-Sept-23

Precision Digital Voltage Meter SN: 2142007

Calibration Date: 2013-Sept-23

Machine re-calibration due date:

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: 

Date: 2013-Oct-18

Signature