

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MB430146G
 Instrument Number: 8455818
 IP Address: 192.168.111.222

Date Tested: 28/08/13
 Mac Address: 00:0A:8D:81:06:8A

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	382	381	882	880	1382	1380
Channel 2	382	381	882	880	1382	1380
Channel 3	382	382	882	882	1382	1382
Channel 4	382	381	882	881	1382	1381
Channel 5	382	381	882	881	1382	1380
Channel 6	382	383	882	882	1382	1381

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055 Calibration Date: 09/08/13
 Precision digital multimeter SN: 2142011 Calibration Date: 09/08/13
 ProHeat 35 power supply used for signal output SN: LK160098G Calibration date: 28/08/13

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.
 Calibration due date: 28/08/14

Technician: Guno Rodin
 Signature

Montreal Canada office