

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MC040342G Date Tested: Nov 29, 2016
 Instrument Number: 129542 Mac Address: 00:0A:8D:01:FA:06
 IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.780	381	6.007	881	9.235	1382
Channel 2	2.775	380	5.997	880	9.236	1381
Channel 3	2.787	382	6.013	882	9.243	1382
Channel 4	2.802	384	6.025	883	9.247	1383
Channel 5	2.799	384	6.025	883	9.245	1383
Channel 6	2.794	383	6.022	883	9.250	1384

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118058 Calibration Date: Sept 27/2016
 Precision digital multimeter SN: 2142007 Calibration Date: Sept 28/2016
 ProHeat 35 power supply used for signal output SN: IH-MME340056 Calibration date: Nov 29, 2016

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.
 Calibration due date:

Technician: [Signature]
 Signature

Nov 29, 2017