

# Red-D-Arc Welderrentals

## Calibration Certificate

Unit Serial Number: MC290381G  
 Instrument Number: 130989  
 IP Address: 192.168.111.222

Date Tested: 23/08/13  
 Mac Address: 00:0A:8D:01:FF:AD

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	382	382	882	882	1382	1383
Channel 2	382	382	882	884	1382	1383
Channel 3	382	382	882	884	1382	1383
Channel 4	382	382	882	883	1382	1384
Channel 5	382	382	882	882	1382	1383
Channel 6	382	382	882	882	1382	1383

**Allowable difference in temp reading is +or- 5 Deg. F.**

**Instruments Used:**

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 09/08/13

Precision digital multimeter SN: 2142011

Calibration Date: 09/08/13

ProHeat 35 power supply used for signal output SN: LK230189G

Calibration date: 23/08/13

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date: 23/08/14

Technician: Guro Rodin

Signature

Montreal Canada office