

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MD200144

Date Tested: 8-18-14

Instrument Number: 8463917

Mac Address: 00:0A:8D:81:26:2D

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.189	383	6.011	882	9.248	1382
Channel 2	2.788	382	6.015	883	9.246	1382
Channel 3	2.791	383	6.015	883	9.246	1382
Channel 4	2.788	382	6.013	882	9.241	1381
Channel 5	2.785	382	6.010	882	9.236	1381
Channel 6	2.791	382	6.011	882	9.101	1383

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1897018

Calibration Date: 8-4-14

Precision digital multimeter SN: US36049893

Calibration Date: 8-4-14

ProHeat 35 power supply used for signal output SN: LK110333G

Calibration date: 8-18-14

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: CHRIS NIETERS

Calibration due date: 8-18-15

Signature